

**Description: Microtox<sup>®</sup> Acute Reagent**

**Part Number:** AZF686018A

**Lot Number:** 12A4000A

**Expiration Date:** January 2014

**Storage Condition:** Frozen, -20 °C

**Quality Control Testing**


Test	Results	Specification	Pass/Fail
Phenol	EC50 (5 min) = 18.0 mg/L	13.0-26.0 mg/L	Pass
Zinc Sulfate	EC50 (15 min) = 4.3 mg/L	3.0-10.0 mg/L	Pass
Zinc ++ (ion) *Calculated	EC50 (15 min) = 0.9 mg/L	0.6-2.2 mg/L	Pass
DeltaTox Adjusted Light Reading	103	≥ 30	Pass

The values above are the mean of all Quality Control results obtained for this lot.

All results are obtained using the Quality Control Procedure with an approved lot of

Microtox 2% Diluent and Reconstitution Solution on the Model 500 and DeltaTox

Analyzers.

Reviewed by:  
  
22Feb12