



MODERNWATER

## Certificate of Analysis

**Description:** Microtox<sup>®</sup> Acute Reagent

**Part Number:** AZF686018A

**Lot Number:** 12B4010

**Expiration Date:** February 2014

**Storage Condition:** Frozen, -20 °C

### Quality Control Testing

Test	Results	Specification	Pass/Fail
Phenol	EC50 (5 min) = 16.2 mg/L	13.0-26.0 mg/L	Pass
Zinc Sulfate	EC50 (15 min) = 3.7 mg/L	3.0-10.0 mg/L	Pass
Zinc ++ (ion) *Calculated	EC50 (15 min) = 0.8 mg/L	0.6-2.2 mg/L	Pass
DeltaTox Adjusted Light Reading	121	≥ 30	Pass

The values above are the mean of all Quality Control results obtained for this lot.

All results are obtained using the Quality Control Procedure with an approved lot of Microtox 2% Diluent and Reconstitution Solution on the Model 500 and DeltaTox Analyzers.

Reviewed by:

  
14 Mar 12

Modern Water Inc., 15 Read's Way, Suite 100, New Castle, DE 19720 USA

Tel: 1-302-669-6900; 1-855-637-6426; [www.modernwater.com](http://www.modernwater.com)